

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071624\
 Data File : VY018830.D
 Acq On : 16 Jul 2024 14:15
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jul 17 01:55:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.771	168	196179	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	343825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	308357	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	133380	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.124	65	94407	48.904	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.800%
35) Dibromofluoromethane	7.704	113	105396	48.943	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.880%
50) Toluene-d8	10.173	98	384972	47.186	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.380%
62) 4-Bromofluorobenzene	12.477	95	129508	46.482	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071624\
Data File : VY018830.D
Acq On : 16 Jul 2024 14:15
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jul 17 01:55:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

