

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071624\
 Data File : VY018833.D
 Acq On : 16 Jul 2024 15:30
 Operator : SY/MD
 Sample : P3246-14
 Misc : 7.28g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-CTW2-BL-03-G

Quant Time: Jul 17 01:57:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.771	168	201604	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.673	114	345960	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	313042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	127125	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.118	65	114733	57.834	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.660%
35) Dibromofluoromethane	7.691	113	110406	50.953	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.900%
50) Toluene-d8	10.167	98	394888	48.103	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.200%
62) 4-Bromofluorobenzene	12.477	95	130010	46.375	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.740%
Target Compounds						Qvalue
16) Acetone	3.918	43	5269	14.406	ug/l #	85
25) 2-Butanone	6.960	43	987	1.925	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071624\
Data File : VY018833.D
Acq On : 16 Jul 2024 15:30
Operator : SY/MD
Sample : P3246-14
Misc : 7.28g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-CTW2-BL-03-G

Quant Time: Jul 17 01:57:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

