

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071723\
 Data File : VY014671.D
 Acq On : 17 Jul 2023 14:44
 Operator : SY/MD
 Sample : 03580-02
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SAMPLE-1

Quant Time: Jul 18 06:09:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	12315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	19441	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	17818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	7358	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	7301	52.026	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.060%
35) Dibromofluoromethane	7.710	113	5459	44.111	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.220%
50) Toluene-d8	10.179	98	22954	50.309	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.620%
62) 4-Bromofluorobenzene	12.477	95	8401	48.020	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071723\
Data File : VY014671.D
Acq On : 17 Jul 2023 14:44
Operator : SY/MD
Sample : 03580-02
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SAMPLE-1

Quant Time: Jul 18 06:09:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

