

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071723\
 Data File : VY014678.D
 Acq On : 17 Jul 2023 17:30
 Operator : SY/MD
 Sample : 03631-04
 Misc : 6.26g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-04(2-4)

Quant Time: Jul 18 06:12:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	211735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	369783	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	350135	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	172137	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	140201	58.107	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 116.220%	
35) Dibromofluoromethane	7.710	113	118095	50.169	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.340%	
50) Toluene-d8	10.179	98	441186	50.837	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.680%	
62) 4-Bromofluorobenzene	12.477	95	169373	50.898	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.800%	
Target Compounds						
3) Chloromethane	2.107	50	2608	1.345	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071723\
Data File : VY014678.D
Acq On : 17 Jul 2023 17:30
Operator : SY/MD
Sample : 03631-04
Misc : 6.26g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-04(2-4)

Quant Time: Jul 18 06:12:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

