

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071823\
 Data File : VY014702.D
 Acq On : 18 Jul 2023 15:20
 Operator : SY/MD
 Sample : 03640-21
 Misc : 5.30g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP12

Quant Time: Jul 19 04:05:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	202596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	355267	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	334726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	156832	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	125651	54.426	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	108.860%	
35) Dibromofluoromethane	7.716	113	112205	49.614	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	99.220%	
50) Toluene-d8	10.179	98	417737	50.101	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	100.200%	
62) 4-Bromofluorobenzene	12.477	95	157759	49.345	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	98.700%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	25657	7.179	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071823\
Data File : VY014702.D
Acq On : 18 Jul 2023 15:20
Operator : SY/MD
Sample : 03640-21
Misc : 5.30g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP12

Quant Time: Jul 19 04:05:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

