

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071823\
 Data File : VY014706.D
 Acq On : 18 Jul 2023 17:08
 Operator : SY/MD
 Sample : 03641-11
 Misc : 4.51g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP14

Quant Time: Jul 19 04:06:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	198596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	343603	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	326931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	154577	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	125659	55.526	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.060%
35) Dibromofluoromethane	7.716	113	109231	49.939	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.880%
50) Toluene-d8	10.179	98	409248	50.749	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.500%
62) 4-Bromofluorobenzene	12.477	95	152780	49.410	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071823\
Data File : VY014706.D
Acq On : 18 Jul 2023 17:08
Operator : SY/MD
Sample : 03641-11
Misc : 4.51g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP14

Quant Time: Jul 19 04:06:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

