

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071823\
 Data File : VY014707.D
 Acq On : 18 Jul 2023 17:31
 Operator : SY/MD
 Sample : 03641-21
 Misc : 3.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP16

Quant Time: Jul 19 04:07:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	196571	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	344605	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	327001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	155135	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	125236	55.909	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.820%
35) Dibromofluoromethane	7.716	113	108820	49.606	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.220%
50) Toluene-d8	10.179	98	408085	50.458	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.920%
62) 4-Bromofluorobenzene	12.477	95	153602	49.532	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071823\
Data File : VY014707.D
Acq On : 18 Jul 2023 17:31
Operator : SY/MD
Sample : 03641-21
Misc : 3.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP16

Quant Time: Jul 19 04:07:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

