

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071823\
 Data File : VY014709.D
 Acq On : 18 Jul 2023 18:18
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 19 04:07:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	65	0.00
2 T	Dichlorodifluoromethane	0.407	0.363	10.8	61	0.00
3 P	Chloromethane	0.458	0.515	-12.4	75	0.00
4 C	Vinyl Chloride	0.512	0.601	-17.4#	76	0.00
5 T	Bromomethane	0.338	0.449	-32.8#	88	0.00
6 T	Chloroethane	0.320	0.387	-20.9	78	0.00
7 T	Trichlorofluoromethane	0.860	0.865	-0.6	64	0.00
8 T	Diethyl Ether	0.320	0.330	-3.1	67	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.512	2.8	62	0.00
10 T	Methyl Iodide	0.493	0.475	3.7	57	0.00
11 T	Tert butyl alcohol	0.125	0.124	0.8	73	0.00
12 CM	1,1-Dichloroethene	0.492	0.479	2.6#	62	0.00
13 T	Acrolein	0.060	0.032	46.7#	39#	0.00
14 T	Allyl chloride	0.874	0.810	7.3	60	0.00
15 T	Acrylonitrile	0.176	0.202	-14.8	74	0.00
16 T	Acetone	0.181	0.171	5.5	57	0.00
17 T	Carbon Disulfide	1.410	1.273	9.7	60	0.00
18 T	Methyl Acetate	0.693	0.804	-16.0	75	0.00
19 T	Methyl tert-butyl Ether	1.599	1.711	-7.0	68	0.00
20 T	Methylene Chloride	0.656	0.671	-2.3	73	0.00
21 T	trans-1,2-Dichloroethene	0.559	0.557	0.4	64	0.00
22 T	Diisopropyl ether	1.777	1.784	-0.4	64	0.00
23 T	Vinyl Acetate	1.064	1.106	-3.9	66	0.00
24 P	1,1-Dichloroethane	1.017	1.024	-0.7	64	0.00
25 T	2-Butanone	0.264	0.285	-8.0	69	0.00
26 T	2,2-Dichloropropane	0.969	0.881	9.1	58	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.681	-4.8	67	0.00
28 T	Bromochloromethane	0.438	0.441	-0.7	69	0.00
29 T	Tetrahydrofuran	0.163	0.187	-14.7	75	0.00
30 C	Chloroform	1.049	1.073	-2.3#	66	0.00
31 T	Cyclohexane	0.946	0.823	13.0	59	0.00
32 T	1,1,1-Trichloroethane	0.944	0.963	-2.0	66	0.00
33 S	1,2-Dichloroethane-d4	0.570	0.577	-1.2	67	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	67	0.00
35 S	Dibromofluoromethane	0.318	0.320	-0.6	66	0.00
36 T	1,1-Dichloropropene	0.476	0.460	3.4	62	0.00
37 T	Ethyl Acetate	0.328	0.362	-10.4	74	0.00
38 T	Carbon Tetrachloride	0.503	0.515	-2.4	65	0.00
39 T	Methylcyclohexane	0.620	0.568	8.4	59	0.00
40 TM	Benzene	1.354	1.360	-0.4	64	0.00
41 T	Methacrylonitrile	0.169	0.165	2.4	65	0.00
42 TM	1,2-Dichloroethane	0.429	0.450	-4.9	67	0.00
43 T	Isopropyl Acetate	0.624	0.668	-7.1	69	0.00
44 TM	Trichloroethene	0.379	0.386	-1.8	66	0.00
45 C	1,2-Dichloropropane	0.353	0.352	0.3#	64	0.00
46 T	Dibromomethane	0.214	0.230	-7.5	69	0.00
47 T	Bromodichloromethane	0.505	0.523	-3.6	66	0.00
48 T	Methyl methacrylate	0.281	0.300	-6.8	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071823\
 Data File : VY014709.D
 Acq On : 18 Jul 2023 18:18
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 19 04:07:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.004	-33.3#	77	0.00
50 S	Toluene-d8	1.173	1.153	1.7	66	0.00
51 T	4-Methyl-2-Pentanone	0.328	0.376	-14.6	74	0.00
52 CM	Toluene	0.866	0.876	-1.2#	65	0.00
53 T	t-1,3-Dichloropropene	0.553	0.552	0.2	64	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.597	0.7	63	0.00
55 T	1,1,2-Trichloroethane	0.299	0.323	-8.0	69	0.00
56 T	Ethyl methacrylate	0.456	0.488	-7.0	68	0.00
57 T	1,3-Dichloropropane	0.515	0.543	-5.4	67	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.213	0.0	66	0.00
59 T	2-Hexanone	0.239	0.273	-14.2	73	0.00
60 T	Dibromochloromethane	0.366	0.395	-7.9	69	0.00
61 T	1,2-Dibromoethane	0.298	0.324	-8.7	70	0.00
62 S	4-Bromofluorobenzene	0.450	0.436	3.1	65	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	67	0.00
64 T	Tetrachloroethene	0.402	0.408	-1.5	67	0.00
65 PM	Chlorobenzene	1.065	1.078	-1.2	66	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.419	-4.2	67	0.00
67 C	Ethyl Benzene	1.924	1.914	0.5#	65	0.00
68 T	m/p-Xylenes	0.734	0.747	-1.8	67	0.00
69 T	o-Xylene	0.715	0.729	-2.0	66	0.00
70 T	Styrene	1.206	1.259	-4.4	67	0.00
71 P	Bromoform	0.272	0.306	-12.5	72	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	0.00
73 T	Isopropylbenzene	3.915	3.788	3.2	67	0.00
74 T	N-amyl acetate	1.311	1.343	-2.4	70	0.00
75 P	1,1,2,2-Tetrachloroethane	0.893	0.951	-6.5	74	0.00
76 T	1,2,3-Trichloropropane	0.631	0.673	-6.7	74	0.00
77 T	Bromobenzene	0.894	0.906	-1.3	69	0.00
78 T	n-propylbenzene	4.718	4.544	3.7	66	0.00
79 T	2-Chlorotoluene	2.717	2.601	4.3	65	0.00
80 T	1,3,5-Trimethylbenzene	3.270	3.178	2.8	66	0.00
81 T	trans-1,4-Dichloro-2-butene	0.326	0.310	4.9	64	0.00
82 T	4-Chlorotoluene	2.799	2.684	4.1	65	0.00
83 T	tert-Butylbenzene	2.900	2.835	2.2	66	0.00
84 T	1,2,4-Trimethylbenzene	3.209	3.138	2.2	67	0.00
85 T	sec-Butylbenzene	4.255	4.146	2.6	66	0.00
86 T	p-Isopropyltoluene	3.474	3.431	1.2	67	0.00
87 T	1,3-Dichlorobenzene	1.739	1.777	-2.2	69	0.00
88 T	1,4-Dichlorobenzene	1.743	1.767	-1.4	69	0.00
89 T	n-Butylbenzene	3.380	3.236	4.3	64	0.00
90 T	Hexachloroethane	0.714	0.683	4.3	65	0.00
91 T	1,2-Dichlorobenzene	1.594	1.629	-2.2	69	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.171	0.180	-5.3	74	0.00
93 T	1,2,4-Trichlorobenzene	1.012	1.024	-1.2	67	0.00
94 T	Hexachlorobutadiene	0.556	0.548	1.4	65	0.00
95 T	Naphthalene	2.357	2.557	-8.5	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY071823\
Data File : VY014709.D
Acq On : 18 Jul 2023 18:18
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 19 04:07:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	0.932	-3.7	69	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6