

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY072020\
 Data File : VY003234.D
 Acq On : 20 Jul 2020 16:12
 Operator : SY/MD
 Sample : L3308-02RE
 Misc : 5.03G/5ML/MSVOA Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 2RE

Quant Time: Jul 21 04:39:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y071020S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 13:18:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	397857	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	604541	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	532687	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	241530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.15	65	220244	65.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	131.18%	
35) Dibromofluoromethane	7.72	113	194576	55.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.86%	
50) Toluene-d8	10.18	98	722787	55.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.76%	
62) 4-Bromofluorobenzene	12.48	95	227110	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
Target Compounds						
20) Methylene Chloride	4.72	84	44239	6.996	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.78	63	4833	2.157	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY072020\
Data File : VY003234.D
Acq On : 20 Jul 2020 16:12
Operator : SY/MD
Sample : L3308-02RE
Misc : 5.03G/5ML/MSVOA Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
2RE

Quant Time: Jul 21 04:39:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y071020S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 13:18:03 2020
Response via : Initial Calibration

