

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072022\
 Data File : VY009613.D
 Acq On : 20 Jul 2022 15:31
 Operator : KP/MD
 Sample : N3778-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-03-071922

Quant Time: Jul 21 02:14:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	232528	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	392183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	366072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	172474	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	104431	47.487	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.980%
35) Dibromofluoromethane	7.709	113	102349	41.424	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.840%
50) Toluene-d8	10.172	98	384440	40.475	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	80.940%
62) 4-Bromofluorobenzene	12.477	95	126857	41.218	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	82.440%
Target Compounds						
2) Dichlorodifluoromethane	1.900	85	10988	5.385	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072022\
Data File : VY009613.D
Acq On : 20 Jul 2022 15:31
Operator : KP/MD
Sample : N3778-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-03-071922

Quant Time: Jul 21 02:14:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

