

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072022\
 Data File : VY009614.D
 Acq On : 20 Jul 2022 15:54
 Operator : KP/MD
 Sample : N3777-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-071922

Quant Time: Jul 21 02:14:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	198836	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	325277	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	317779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	149175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	87857	46.720	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.440%
35) Dibromofluoromethane	7.710	113	85769	41.854	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.700%
50) Toluene-d8	10.173	98	332754	42.239	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	84.480%
62) 4-Bromofluorobenzene	12.477	95	109891	43.050	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	86.100%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.900	85	5736	3.288	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072022\
Data File : VY009614.D
Acq On : 20 Jul 2022 15:54
Operator : KP/MD
Sample : N3777-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-071922

Quant Time: Jul 21 02:14:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

