

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072022\
 Data File : VY009616.D
 Acq On : 20 Jul 2022 16:42
 Operator : KP/MD
 Sample : N3726-02
 Misc : 6.31g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OWS-12-12.5

Quant Time: Jul 21 02:14:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	21451	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	31458	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	27637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	10673	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	8282	40.823	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.640%
35) Dibromofluoromethane	7.704	113	5988	30.214	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	60.420%
50) Toluene-d8	10.173	98	29310	38.471	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	76.940%#
62) 4-Bromofluorobenzene	12.477	95	7639	30.943	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	61.880%
Target Compounds						
					Qvalue	
16) Acetone	3.936	43	1121	21.843	ug/l #	54
20) Methylene Chloride	4.698	84	4356	8.764	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072022\
Data File : VY009616.D
Acq On : 20 Jul 2022 16:42
Operator : KP/MD
Sample : N3726-02
Misc : 6.31g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OWS-12-12.5

Quant Time: Jul 21 02:14:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

