

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072023\
 Data File : VY014726.D
 Acq On : 20 Jul 2023 17:15
 Operator : SY/MD
 Sample : 03674-03
 Misc : 6.18g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WO-9

Quant Time: Jul 21 01:33:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	215926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	352600	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	301440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	117349	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	112904	45.885	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.780%
35) Dibromofluoromethane	7.716	113	105707	47.095	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.180%
50) Toluene-d8	10.179	98	408628	49.380	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.760%
62) 4-Bromofluorobenzene	12.477	95	133338	42.022	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.040%
Target Compounds						
					Qvalue	
64) Tetrachloroethene	10.715	164	8726	3.599	ug/l	95
95) Naphthalene	15.233	128	35202	6.364	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072023\
Data File : VY014726.D
Acq On : 20 Jul 2023 17:15
Operator : SY/MD
Sample : 03674-03
Misc : 6.18g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WO-9

Quant Time: Jul 21 01:33:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

