

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072023\
 Data File : VY014727.D
 Acq On : 20 Jul 2023 17:39
 Operator : SY/MD
 Sample : 03681-02
 Misc : 4.97g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RPTP-1-GRAB-3

Quant Time: Jul 21 01:33:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	82043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	126462	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	103727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	44236	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	34568	36.975	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	73.940%		
35) Dibromofluoromethane	7.710	113	36042	44.771	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	89.540%		
50) Toluene-d8	10.179	98	145673	49.082	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	98.160%		
62) 4-Bromofluorobenzene	12.477	95	46367	40.743	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	81.480%		
Target Compounds						
16) Acetone	3.942	43	27436	92.607	ug/l	99
78) n-propylbenzene	12.666	91	4396	1.053	ug/l	96
95) Naphthalene	15.233	128	22761	10.916	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072023\
Data File : VY014727.D
Acq On : 20 Jul 2023 17:39
Operator : SY/MD
Sample : 03681-02
Misc : 4.97g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RPTP-1-GRAB-3

Quant Time: Jul 21 01:33:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

