

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072122\
 Data File : VY009637.D
 Acq On : 21 Jul 2022 16:48
 Operator : KP/MD
 Sample : N3786-15
 Misc : 6.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 0718-WC-TP-WC5B

Quant Time: Jul 22 02:30:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	217925	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	355967	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	331563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	156420	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	118264	57.381	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.760%
35) Dibromofluoromethane	7.710	113	111541	49.738	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.480%
50) Toluene-d8	10.173	98	406311	47.130	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	94.260%
62) 4-Bromofluorobenzene	12.477	95	131631	47.120	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	94.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072122\
Data File : VY009637.D
Acq On : 21 Jul 2022 16:48
Operator : KP/MD
Sample : N3786-15
Misc : 6.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
0718-WC-TP-WC5B

Quant Time: Jul 22 02:30:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

