

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072122\
 Data File : VY009639.D
 Acq On : 21 Jul 2022 17:36
 Operator : KP/MD
 Sample : N3786-19
 Misc : 7.42g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 0718-WC-TP-WC17B

Quant Time: Jul 22 02:30:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	252049	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	407832	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	370569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	176181	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	130830	54.884	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.760%
35) Dibromofluoromethane	7.710	113	124814	48.578	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.160%
50) Toluene-d8	10.173	98	461160	46.689	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	93.380%
62) 4-Bromofluorobenzene	12.477	95	149542	46.724	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	93.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072122\
Data File : VY009639.D
Acq On : 21 Jul 2022 17:36
Operator : KP/MD
Sample : N3786-19
Misc : 7.42g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
0718-WC-TP-WC17B

Quant Time: Jul 22 02:30:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

