

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072122\
 Data File : VY009644.D
 Acq On : 21 Jul 2022 19:35
 Operator : KP/MD
 Sample : N3835-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-191E

Quant Time: Jul 22 02:31:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	269924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	446151	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	423770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	197567	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	143095	56.054	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.100%
35) Dibromofluoromethane	7.704	113	139493	49.629	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.260%
50) Toluene-d8	10.173	98	513684	47.540	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	95.080%
62) 4-Bromofluorobenzene	12.477	95	167395	47.810	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	95.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072122\
Data File : VY009644.D
Acq On : 21 Jul 2022 19:35
Operator : KP/MD
Sample : N3835-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-191E

Quant Time: Jul 22 02:31:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

