

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
 Data File : VY014743.D
 Acq On : 21 Jul 2023 13:45
 Operator : SY/MD
 Sample : 03681-11
 Misc : 4.60g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RPTP-4-GRAB-7

Quant Time: Jul 22 05:36:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	215538	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	376281	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	355055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	167664	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	136097	55.411	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.820%
35) Dibromofluoromethane	7.710	113	121734	50.822	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.640%
50) Toluene-d8	10.173	98	444065	50.285	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.560%
62) 4-Bromofluorobenzene	12.477	95	166340	49.124	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
Data File : VY014743.D
Acq On : 21 Jul 2023 13:45
Operator : SY/MD
Sample : 03681-11
Misc : 4.60g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RPTP-4-GRAB-7

Quant Time: Jul 22 05:36:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

