

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
 Data File : VY014744.D
 Acq On : 21 Jul 2023 14:08
 Operator : SY/MD
 Sample : 03681-12
 Misc : 5.11g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RPTP-4-GRAB-8

Quant Time: Jul 22 05:36:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	214660	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	369620	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	349811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	163461	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	132066	53.990	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 107.980%	
35) Dibromofluoromethane	7.710	113	117861	50.092	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.180%	
50) Toluene-d8	10.179	98	440168	50.742	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.480%	
62) 4-Bromofluorobenzene	12.477	95	166138	49.948	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 99.900%	
Target Compounds						Qvalue
16) Acetone	3.942	43	122899	158.549	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
Data File : VY014744.D
Acq On : 21 Jul 2023 14:08
Operator : SY/MD
Sample : 03681-12
Misc : 5.11g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RPTP-4-GRAB-8

Quant Time: Jul 22 05:36:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

