

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
 Data File : VY014746.D
 Acq On : 21 Jul 2023 14:56
 Operator : SY/MD
 Sample : 03681-15
 Misc : 4.88g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RPTP-5-GRAB-10

Quant Time: Jul 22 05:37:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	210810	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	365988	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	345697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	160930	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	129173	53.771	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.540%
35) Dibromofluoromethane	7.716	113	115096	49.402	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.800%
50) Toluene-d8	10.179	98	432484	50.351	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.700%
62) 4-Bromofluorobenzene	12.477	95	161662	49.085	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
Data File : VY014746.D
Acq On : 21 Jul 2023 14:56
Operator : SY/MD
Sample : 03681-15
Misc : 4.88g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RPTP-5-GRAB-10

Quant Time: Jul 22 05:37:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

