

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
 Data File : VY014752.D
 Acq On : 21 Jul 2023 17:18
 Operator : SY/MD
 Sample : 03706-02
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P9B

Quant Time: Jul 22 05:39:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	2939	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	4955	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	4451	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1627	48.580	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.160%
35) Dibromofluoromethane	7.716	113	684	21.685	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	43.380%#
50) Toluene-d8	10.179	98	5764	49.566	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.140%
62) 4-Bromofluorobenzene	12.477	95	1678	37.632	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.260%
Target Compounds						
16) Acetone	3.942	43	1299	122.398	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
Data File : VY014752.D
Acq On : 21 Jul 2023 17:18
Operator : SY/MD
Sample : 03706-02
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P9B

Quant Time: Jul 22 05:39:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

