

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
 Data File : VY014753.D
 Acq On : 21 Jul 2023 17:42
 Operator : SY/MD
 Sample : 03706-03
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P10A

Quant Time: Jul 22 05:40:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	6011	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	10431	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	9365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3200	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	3894	56.849	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.700%
35) Dibromofluoromethane	7.722	113	1671	25.165	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	50.340%#
50) Toluene-d8	10.179	98	11758	48.030	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.060%
62) 4-Bromofluorobenzene	12.483	95	3896	41.505	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.000%
Target Compounds						
						Qvalue
3) Chloromethane	2.107	50	574	10.427	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
Data File : VY014753.D
Acq On : 21 Jul 2023 17:42
Operator : SY/MD
Sample : 03706-03
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P10A

Quant Time: Jul 22 05:40:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

