

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
 Data File : VY014756.D
 Acq On : 21 Jul 2023 18:53
 Operator : SY/MD
 Sample : 03706-06
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P20B

Quant Time: Jul 22 05:41:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	12034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	19982	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	17484	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	6439	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	6984	50.929	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.860%
35) Dibromofluoromethane	7.716	113	4880	38.365	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.720%
50) Toluene-d8	10.179	98	22676	48.354	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.700%
62) 4-Bromofluorobenzene	12.477	95	7466	41.520	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.040%
Target Compounds						
3) Chloromethane	2.113	50	364	3.303	ug/l	93
16) Acetone	3.954	43	1196	27.522	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
Data File : VY014756.D
Acq On : 21 Jul 2023 18:53
Operator : SY/MD
Sample : 03706-06
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P20B

Quant Time: Jul 22 05:41:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

