

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
 Data File : VY014758.D
 Acq On : 21 Jul 2023 19:40
 Operator : SY/MD
 Sample : 03706-08
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P21B

Quant Time: Jul 22 05:41:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	1003	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	1612	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	132	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	743	65.007	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.020%
35) Dibromofluoromethane	7.722	113	210	20.465	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	40.920%#
50) Toluene-d8	10.185	98	1992	52.653	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.300%
62) 4-Bromofluorobenzene	12.483	95	199	13.718	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	27.440%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072123\
Data File : VY014758.D
Acq On : 21 Jul 2023 19:40
Operator : SY/MD
Sample : 03706-08
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P21B

Quant Time: Jul 22 05:41:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

