

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072221\
 Data File : VY005415.D
 Acq On : 22 Jul 2021 16:41
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY072221

Quant Time: Jul 23 06:25:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.453	0.426	6.0	121	0.00
3 P	Chloromethane	0.555	0.549	1.1	122	0.00
4 C	Vinyl Chloride	0.661	0.647	2.1#	121	0.00
5 T	Bromomethane	0.483	0.500	-3.5	113	0.00
6 T	Chloroethane	0.415	0.437	-5.3	117	0.00
7 T	Trichlorofluoromethane	0.864	0.883	-2.2	115	0.00
8 T	Diethyl Ether	0.304	0.324	-6.6	118	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.470	0.495	-5.3	122	0.00
10 T	Methyl Iodide	0.533	0.602	-12.9	129	0.00
11 T	Tert butyl alcohol	0.050	0.050	0.0	126	-0.02
12 CM	1,1-Dichloroethene	0.478	0.505	-5.6#	120	0.00
13 T	Acrolein	0.046	0.041	10.9	123	0.00
14 T	Allyl chloride	0.855	0.904	-5.7	119	0.00
15 T	Acrylonitrile	0.151	0.159	-5.3	120	0.00
16 T	Acetone	0.122	0.126	-3.3	134	0.00
17 T	Carbon Disulfide	1.581	1.647	-4.2	120	0.00
18 T	Methyl Acetate	0.343	0.353	-2.9	121	0.00
19 T	Methyl tert-butyl Ether	1.442	1.583	-9.8	118	0.00
20 T	Methylene Chloride	0.707	0.709	-0.3	123	0.00
21 T	trans-1,2-Dichloroethene	0.537	0.572	-6.5	118	0.00
22 T	Diisopropyl ether	1.660	1.641	1.1	104	0.00
23 T	Vinyl Acetate	1.092	1.197	-9.6	113	0.00
24 P	1,1-Dichloroethane	0.915	1.022	-11.7	118	0.00
25 T	2-Butanone	0.183	0.175	4.4	108	0.00
26 T	2,2-Dichloropropane	0.832	0.811	2.5	106	0.00
27 T	cis-1,2-Dichloroethene	0.574	0.561	2.3	103	0.00
28 T	Bromochloromethane	0.368	0.332	9.8	99	0.00
29 T	Tetrahydrofuran	0.121	0.119	1.7	106	0.00
30 C	Chloroform	0.905	0.904	0.1#	101	0.00
31 T	Cyclohexane	0.912	0.863	5.4	106	0.00
32 T	1,1,1-Trichloroethane	0.838	0.839	-0.1	103	0.00
33 S	1,2-Dichloroethane-d4	0.534	0.518	3.0	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.289	0.286	1.0	97	0.00
36 T	1,1-Dichloropropene	0.483	0.490	-1.4	104	0.00
37 T	Ethyl Acetate	0.294	0.292	0.7	102	0.00
38 T	Carbon Tetrachloride	0.528	0.534	-1.1	104	0.00
39 T	Methylcyclohexane	0.620	0.617	0.5	104	0.00
40 TM	Benzene	1.389	1.402	-0.9	102	0.00
41 T	Methacrylonitrile	0.154	0.169	-9.7	107	0.00
42 TM	1,2-Dichloroethane	0.449	0.456	-1.6	102	0.00
43 T	Isopropyl Acetate	0.533	0.549	-3.0	103	0.00
44 TM	Trichloroethene	0.377	0.380	-0.8	103	0.00
45 C	1,2-Dichloropropane	0.342	0.343	-0.3#	102	0.00
46 T	Dibromomethane	0.188	0.191	-1.6	102	0.00
47 T	Bromodichloromethane	0.473	0.489	-3.4	102	0.00
48 T	Methyl methacrylate	0.248	0.251	-1.2	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072221\
 Data File : VY005415.D
 Acq On : 22 Jul 2021 16:41
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY072221

Quant Time: Jul 23 06:25:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	105	-0.02
50 S	Toluene-d8	1.224	1.209	1.2	97	0.00
51 T	4-Methyl-2-Pentanone	0.283	0.290	-2.5	104	0.00
52 CM	Toluene	0.882	0.892	-1.1#	102	0.00
53 T	t-1,3-Dichloropropene	0.529	0.543	-2.6	102	0.00
54 T	cis-1,3-Dichloropropene	0.582	0.595	-2.2	103	0.00
55 T	1,1,2-Trichloroethane	0.270	0.275	-1.9	101	0.00
56 T	Ethyl methacrylate	0.408	0.425	-4.2	102	0.00
57 T	1,3-Dichloropropane	0.481	0.489	-1.7	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.195	0.211	-8.2	106	0.00
59 T	2-Hexanone	0.198	0.205	-3.5	105	0.00
60 T	Dibromochloromethane	0.330	0.339	-2.7	101	0.00
61 T	1,2-Dibromoethane	0.256	0.263	-2.7	102	0.00
62 S	4-Bromofluorobenzene	0.404	0.393	2.7	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.378	0.371	1.9	102	0.00
65 PM	Chlorobenzene	1.024	1.022	0.2	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.382	0.387	-1.3	102	0.00
67 C	Ethyl Benzene	1.864	1.874	-0.5#	102	0.00
68 T	m/p-Xylenes	0.724	0.728	-0.6	103	0.00
69 T	o-Xylene	0.689	0.688	0.1	101	0.00
70 T	Styrene	1.169	1.190	-1.8	101	0.00
71 P	Bromoform	0.244	0.249	-2.0	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73 T	Isopropylbenzene	3.448	3.574	-3.7	103	0.00
74 T	N-amyl acetate	1.050	1.106	-5.3	103	0.00
75 P	1,1,2,2-Tetrachloroethane	0.658	0.684	-4.0	103	0.00
76 T	1,2,3-Trichloropropane	0.492	0.510	-3.7	105	0.00
77 T	Bromobenzene	0.844	0.868	-2.8	101	0.00
78 T	n-propylbenzene	4.160	4.299	-3.3	104	0.00
79 T	2-Chlorotoluene	2.337	2.392	-2.4	102	0.00
80 T	1,3,5-Trimethylbenzene	2.880	2.954	-2.6	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.253	0.259	-2.4	104	0.00
82 T	4-Chlorotoluene	2.418	2.472	-2.2	103	0.00
83 T	tert-Butylbenzene	2.565	2.661	-3.7	104	0.00
84 T	1,2,4-Trimethylbenzene	2.889	2.962	-2.5	102	0.00
85 T	sec-Butylbenzene	3.700	3.844	-3.9	105	0.00
86 T	p-Isopropyltoluene	3.186	3.288	-3.2	105	0.00
87 T	1,3-Dichlorobenzene	1.656	1.706	-3.0	102	0.00
88 T	1,4-Dichlorobenzene	1.641	1.680	-2.4	104	0.00
89 T	n-Butylbenzene	2.924	3.000	-2.6	105	0.00
90 T	Hexachloroethane	0.573	0.591	-3.1	103	0.00
91 T	1,2-Dichlorobenzene	1.479	1.506	-1.8	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.128	0.132	-3.1	107	0.00
93 T	1,2,4-Trichlorobenzene	1.020	1.028	-0.8	104	0.00
94 T	Hexachlorobutadiene	0.647	0.650	-0.5	104	0.00
95 T	Naphthalene	2.052	2.123	-3.5	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072221\
Data File : VY005415.D
Acq On : 22 Jul 2021 16:41
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY072221

Quant Time: Jul 23 06:25:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:20:14 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.920	0.931	-1.2	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6