

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072222\
 Data File : VY009656.D
 Acq On : 22 Jul 2022 13:47
 Operator : KP/MD
 Sample : N3773-14
 Misc : 5.85g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DL-2(7-7.5)

Quant Time: Jul 23 00:53:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	154295	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	249821	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	241775	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	117491	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	82812	56.750	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.500%
35) Dibromofluoromethane	7.710	113	79141	50.285	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.560%
50) Toluene-d8	10.173	98	292818	48.397	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	96.800%
62) 4-Bromofluorobenzene	12.477	95	99052	50.524	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	101.040%
Target Compounds						
						Qvalue
16) Acetone	3.936	43	12555	34.011	ug/l #	80
20) Methylene Chloride	4.704	84	40861	15.204	ug/l #	83
84) 1,2,4-Trimethylbenzene	13.117	105	10784	1.604	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072222\
Data File : VY009656.D
Acq On : 22 Jul 2022 13:47
Operator : KP/MD
Sample : N3773-14
Misc : 5.85g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DL-2(7-7.5)

Quant Time: Jul 23 00:53:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

