

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072222\
 Data File : VY009657.D
 Acq On : 22 Jul 2022 14:13
 Operator : KP/MD
 Sample : N3773-15
 Misc : 5.84g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DL-2(14.5-15)

Quant Time: Jul 23 00:53:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	152019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	243078	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	232428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	109640	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	80270	55.831	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.660%
35) Dibromofluoromethane	7.710	113	79359	51.822	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.640%
50) Toluene-d8	10.173	98	282132	47.924	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	95.840%
62) 4-Bromofluorobenzene	12.477	95	89410	46.871	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	93.740%
Target Compounds						Qvalue
20) Methylene Chloride	4.704	84	41502	16.057	ug/l	# 81
22) Diisopropyl ether	6.106	45	11239	2.434	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072222\
Data File : VY009657.D
Acq On : 22 Jul 2022 14:13
Operator : KP/MD
Sample : N3773-15
Misc : 5.84g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DL-2(14.5-15)

Quant Time: Jul 23 00:53:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

