

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY072320\
 Data File : VY003314.D
 Acq On : 23 Jul 2020 14:22
 Operator : SY/MD
 Sample : L3380-03
 Misc : 6.94G/5ML/MSVOA Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-241

Quant Time: Jul 24 05:24:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y071020S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 13:18:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.80	168	421539	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.69	114	649774	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.49	117	521822	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.42	152	159378	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.14	65	204738	57.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.08%	
35) Dibromofluoromethane	7.72	113	201726	53.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.90%	
50) Toluene-d8	10.18	98	748975	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
62) 4-Bromofluorobenzene	12.48	95	187438	37.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.64%	
Target Compounds						
16) Acetone	3.95	43	5878	7.396	ug/l	98
20) Methylene Chloride	4.72	84	59425	10.522	ug/l	95
84) 1,2,4-Trimethylbenzene	13.12	105	9760	1.146	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY072320\
Data File : VY003314.D
Acq On : 23 Jul 2020 14:22
Operator : SY/MD
Sample : L3380-03
Misc : 6.94G/5ML/MSVOA Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-241

Quant Time: Jul 24 05:24:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y071020S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 13:18:03 2020
Response via : Initial Calibration

