

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
 Data File : VY014768.D
 Acq On : 24 Jul 2023 13:32
 Operator : SY/MD
 Sample : 03706-08
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P21B

Quant Time: Jul 24 14:15:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	16779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	26443	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	21271	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	8559	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	8547	44.701	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.400%
35) Dibromofluoromethane	7.716	113	7410	44.021	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.040%
50) Toluene-d8	10.179	98	29914	48.202	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.400%
62) 4-Bromofluorobenzene	12.483	95	9463	39.767	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
Data File : VY014768.D
Acq On : 24 Jul 2023 13:32
Operator : SY/MD
Sample : 03706-08
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P21B

Quant Time: Jul 24 14:15:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

