

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
 Data File : VY014782.D
 Acq On : 24 Jul 2023 19:04
 Operator : SY/MD
 Sample : 03719-06
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P23B

Quant Time: Jul 25 01:32:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	47952	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	79334	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	69030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	30585	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	25999	47.580	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.160%
35) Dibromofluoromethane	7.716	113	24430	48.374	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.740%
50) Toluene-d8	10.179	98	91163	48.962	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.920%
62) 4-Bromofluorobenzene	12.477	95	31053	43.496	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
Data File : VY014782.D
Acq On : 24 Jul 2023 19:04
Operator : SY/MD
Sample : 03719-06
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P23B

Quant Time: Jul 25 01:32:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

