

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
 Data File : VY014785.D
 Acq On : 24 Jul 2023 20:15
 Operator : SY/MD
 Sample : 03694-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: Jul 25 01:33:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	4000	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	7645	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	7665	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	3115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	3780	82.928	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 165.860%#	
35) Dibromofluoromethane	7.722	113	1523	31.295	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 62.580%	
50) Toluene-d8	10.178	98	9400	52.390	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 104.780%	
62) 4-Bromofluorobenzene	12.483	95	3733	54.261	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 108.520%	
Target Compounds						
					Qvalue	
3) Chloromethane	2.113	50	458	12.502	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
Data File : VY014785.D
Acq On : 24 Jul 2023 20:15
Operator : SY/MD
Sample : 03694-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: Jul 25 01:33:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

