

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
 Data File : VY014786.D
 Acq On : 24 Jul 2023 20:38
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 25 01:33:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	61	0.00
2 T	Dichlorodifluoromethane	0.407	0.302	25.8#	47#	0.00
3 P	Chloromethane	0.458	0.475	-3.7	64	0.00
4 C	Vinyl Chloride	0.512	0.535	-4.5#	63	0.00
5 T	Bromomethane	0.338	0.380	-12.4	69	0.00
6 T	Chloroethane	0.320	0.357	-11.6	67	0.00
7 T	Trichlorofluoromethane	0.860	0.768	10.7	53	0.00
8 T	Diethyl Ether	0.320	0.309	3.4	58	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.449	14.8	50	0.00
10 T	Methyl Iodide	0.493	0.498	-1.0	56	0.00
11 T	Tert butyl alcohol	0.125	0.121	3.2	66	0.00
12 CM	1,1-Dichloroethene	0.492	0.432	12.2#	52	0.00
13 T	Acrolein	0.060	0.045	25.0	51	0.00
14 T	Allyl chloride	0.874	0.730	16.5	50	0.00
15 T	Acrylonitrile	0.176	0.201	-14.2	68	0.00
16 T	Acetone	0.181	0.167	7.7	52	0.00
17 T	Carbon Disulfide	1.410	1.132	19.7	49#	0.00
18 T	Methyl Acetate	0.693	0.852	-22.9	74	0.00
19 T	Methyl tert-butyl Ether	1.599	1.623	-1.5	60	0.00
20 T	Methylene Chloride	0.656	0.641	2.3	65	0.00
21 T	trans-1,2-Dichloroethene	0.559	0.508	9.1	55	0.00
22 T	Diisopropyl ether	1.777	1.684	5.2	56	0.00
23 T	Vinyl Acetate	1.064	1.010	5.1	56	0.00
24 P	1,1-Dichloroethane	1.017	0.944	7.2	55	0.00
25 T	2-Butanone	0.264	0.287	-8.7	65	0.00
26 T	2,2-Dichloropropane	0.969	0.735	24.1	45#	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.624	4.0	57	0.00
28 T	Bromochloromethane	0.438	0.400	8.7	58	0.00
29 T	Tetrahydrofuran	0.163	0.192	-17.8	72	0.00
30 C	Chloroform	1.049	1.009	3.8#	57	0.00
31 T	Cyclohexane	0.946	0.713	24.6	47#	0.00
32 T	1,1,1-Trichloroethane	0.944	0.875	7.3	55	0.00
33 S	1,2-Dichloroethane-d4	0.570	0.596	-4.6	65	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	64	0.00
35 S	Dibromofluoromethane	0.318	0.314	1.3	62	0.00
36 T	1,1-Dichloropropene	0.476	0.409	14.1	53	0.00
37 T	Ethyl Acetate	0.328	0.348	-6.1	69	0.00
38 T	Carbon Tetrachloride	0.503	0.442	12.1	54	0.00
39 T	Methylcyclohexane	0.620	0.484	21.9	48#	0.00
40 TM	Benzene	1.354	1.224	9.6	56	0.00
41 T	Methacrylonitrile	0.169	0.160	5.3	61	0.00
42 TM	1,2-Dichloroethane	0.429	0.414	3.5	59	0.00
43 T	Isopropyl Acetate	0.624	0.626	-0.3	62	0.00
44 TM	Trichloroethene	0.379	0.341	10.0	56	0.00
45 C	1,2-Dichloropropane	0.353	0.319	9.6#	56	0.00
46 T	Dibromomethane	0.214	0.212	0.9	61	0.00
47 T	Bromodichloromethane	0.505	0.474	6.1	58	0.00
48 T	Methyl methacrylate	0.281	0.286	-1.8	63	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
 Data File : VY014786.D
 Acq On : 24 Jul 2023 20:38
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 25 01:33:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.004	-33.3#	74	0.00
50 S	Toluene-d8	1.173	1.127	3.9	62	0.00
51 T	4-Methyl-2-Pentanone	0.328	0.370	-12.8	70	0.00
52 CM	Toluene	0.866	0.797	8.0#	57	0.00
53 T	t-1,3-Dichloropropene	0.553	0.490	11.4	55	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.525	12.6	54	0.00
55 T	1,1,2-Trichloroethane	0.299	0.299	0.0	62	0.00
56 T	Ethyl methacrylate	0.456	0.458	-0.4	61	0.00
57 T	1,3-Dichloropropane	0.515	0.505	1.9	60	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.227	-6.6	68	0.00
59 T	2-Hexanone	0.239	0.270	-13.0	70	0.00
60 T	Dibromochloromethane	0.366	0.364	0.5	61	0.00
61 T	1,2-Dibromoethane	0.298	0.299	-0.3	62	0.00
62 S	4-Bromofluorobenzene	0.450	0.421	6.4	61	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	66	0.00
64 T	Tetrachloroethene	0.402	0.398	1.0	64	0.00
65 PM	Chlorobenzene	1.065	0.969	9.0	58	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.377	6.2	59	0.00
67 C	Ethyl Benzene	1.924	1.695	11.9#	56	0.00
68 T	m/p-Xylenes	0.734	0.653	11.0	57	0.00
69 T	o-Xylene	0.715	0.645	9.8	57	0.00
70 T	Styrene	1.206	1.096	9.1	57	0.00
71 P	Bromoform	0.272	0.275	-1.1	63	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
73 T	Isopropylbenzene	3.915	3.309	15.5	56	0.00
74 T	N-amyl acetate	1.311	1.194	8.9	60	0.00
75 P	1,1,2,2-Tetrachloroethane	0.893	0.856	4.1	65	0.00
76 T	1,2,3-Trichloropropane	0.631	0.623	1.3	66	0.00
77 T	Bromobenzene	0.894	0.810	9.4	60	0.00
78 T	n-propylbenzene	4.718	3.962	16.0	55	0.00
79 T	2-Chlorotoluene	2.717	2.288	15.8	56	0.00
80 T	1,3,5-Trimethylbenzene	3.270	2.804	14.3	56	0.00
81 T	trans-1,4-Dichloro-2-butene	0.326	0.273	16.3	55	0.00
82 T	4-Chlorotoluene	2.799	2.364	15.5	56	0.00
83 T	tert-Butylbenzene	2.900	2.458	15.2	56	0.00
84 T	1,2,4-Trimethylbenzene	3.209	2.772	13.6	57	0.00
85 T	sec-Butylbenzene	4.255	3.591	15.6	56	0.00
86 T	p-Isopropyltoluene	3.474	2.972	14.5	56	0.00
87 T	1,3-Dichlorobenzene	1.739	1.573	9.5	59	0.00
88 T	1,4-Dichlorobenzene	1.743	1.562	10.4	59	0.00
89 T	n-Butylbenzene	3.380	2.786	17.6	54	0.00
90 T	Hexachloroethane	0.714	0.606	15.1	56	0.00
91 T	1,2-Dichlorobenzene	1.594	1.475	7.5	61	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.171	0.170	0.6	67	0.00
93 T	1,2,4-Trichlorobenzene	1.012	0.918	9.3	58	0.00
94 T	Hexachlorobutadiene	0.556	0.471	15.3	54	0.00
95 T	Naphthalene	2.357	2.392	-1.5	65	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
Data File : VY014786.D
Acq On : 24 Jul 2023 20:38
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 25 01:33:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.899	0.841	6.5	60	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6