

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
 Data File : VY014763.D
 Acq On : 24 Jul 2023 11:35
 Operator : SY/MD
 Sample : VY0724SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0724SBL01

Quant Time: Jul 24 13:00:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	203591	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	346941	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	323048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	153475	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	121772	52.488	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.980%
35) Dibromofluoromethane	7.716	113	110175	49.886	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.780%
50) Toluene-d8	10.179	98	408685	50.192	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.380%
62) 4-Bromofluorobenzene	12.477	95	150668	48.258	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.520%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
Data File : VY014763.D
Acq On : 24 Jul 2023 11:35
Operator : SY/MD
Sample : VY0724SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0724SBL01

Quant Time: Jul 24 13:00:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY014763.D
Acq: 24 Jul 2023 11:35




Instrument :
MSVOA_Y
ClientSampleId :
VY0724SBL01


Tgt Ion:168 Resp: 203591
Ion Ratio Lower Upper
168 100
99 54.9 43.4 65.0



#33
1,2-Dichloroethane-d4
Concen: 52.488 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY014763.D
Acq: 24 Jul 2023 11:35

Tgt Ion: 65 Resp: 121772
Ion Ratio Lower Upper
65 100
67 52.3 0.0 101.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.685 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VY014763.D
Acq: 24 Jul 2023 11:35




Instrument :
MSVOA_Y
ClientSampleId :
VY0724SBL01


Tgt Ion:114 Resp: 346941
Ion Ratio Lower Upper
114 100
63 20.1 0.0 42.4
88 15.7 0.0 30.6



#35
Dibromofluoromethane
Concen: 49.886 ug/l
RT: 7.716 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY014763.D
Acq: 24 Jul 2023 11:35

Tgt Ion:113 Resp: 110175
Ion Ratio Lower Upper
113 100
111 102.4 81.3 121.9
192 19.7 16.9 25.3





#50

Toluene-d8

Concen: 50.192 ug/l

RT: 10.179 min Scan# 1371

Delta R.T. -0.000 min

Lab File: VY014763.D

Acq: 24 Jul 2023 11:35

Instrument :

MSVOA_Y

ClientSampleId :





VY0724SBL01

Tgt Ion: 98 Resp: 408685

Ion Ratio Lower Upper

98 100

100 65.7 50.1 75.1



#62

4-Bromofluorobenzene

Concen: 48.258 ug/l

RT: 12.477 min Scan# 1748

Delta R.T. 0.000 min

Lab File: VY014763.D

Acq: 24 Jul 2023 11:35




Tgt Ion: 95 Resp: 150668


Ion Ratio Lower Upper

95 100

174 81.8 0.0 178.2

176 79.9 0.0 173.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 11

Delta R.T. 0.001 min

Lab File: VY014763.D

Acq: 24 Jul 2023 11:35

Instrument :

MSVOA_Y

ClientSampleId :

VY0724SBL01





Tgt Ion:117 Resp: 323048

Ion Ratio Lower Upper

117 100

82 54.7 43.4 65.0

119 32.4 26.3 39.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. 0.000 min

Lab File: VY014763.D

Acq: 24 Jul 2023 11:35

Tgt Ion:152 Resp: 153475

Ion Ratio Lower Upper

152 100

115 59.5 28.8 86.5

150 157.5 0.0 348.8

