



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072423\
Data File : VY014761.D
Acq On : 24 Jul 2023 08:54
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Title : SW846 8260
Last Update : Fri Jul 14 02:09:50 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	21915	PASS
75	95	30	60	52.4	60997	PASS
95	95	100	100	100.0	116499	PASS
96	95	5	9	6.5	7611	PASS
173	174	0.00	2	0.9	854	PASS
174	95	50	100	83.4	97115	PASS
175	174	5	9	7.2	7033	PASS
176	174	95	101	95.3	92512	PASS
177	176	5	9	6.7	6213	PASS