

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072522\
 Data File : VY009680.D
 Acq On : 25 Jul 2022 12:40
 Operator : KP/MD
 Sample : N3858-03
 Misc : 3.19g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-B

Quant Time: Jul 26 05:56:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 04:22:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	172291	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	272705	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	264567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	131968	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	88067	54.047	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.100%	
35) Dibromofluoromethane	7.710	113	85285	49.641	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.280%	
50) Toluene-d8	10.173	98	314229	47.577	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	= 95.160%	
62) 4-Bromofluorobenzene	12.477	95	112625	52.626	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	= 105.260%	
Target Compounds						
2) Dichlorodifluoromethane	1.900	85	29874	19.761	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072522\
Data File : VY009680.D
Acq On : 25 Jul 2022 12:40
Operator : KP/MD
Sample : N3858-03
Misc : 3.19g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-B

Quant Time: Jul 26 05:56:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 04:22:09 2022
Response via : Initial Calibration

