

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
 Data File : VY014789.D
 Acq On : 25 Jul 2023 10:13
 Operator : SY/MD
 Sample : VY0725SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0725SBL01

Quant Time: Jul 25 14:12:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	175842	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	304719	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	287464	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	136557	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	106505	53.152	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.300%
35) Dibromofluoromethane	7.716	113	96298	49.644	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.280%
50) Toluene-d8	10.179	98	362762	50.725	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.477	95	133279	48.604	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
Data File : VY014789.D
Acq On : 25 Jul 2023 10:13
Operator : SY/MD
Sample : VY0725SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0725SBL01

Quant Time: Jul 25 14:12:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

