

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
 Data File : VY014792.D
 Acq On : 25 Jul 2023 11:24
 Operator : SY/MD
 Sample : 03719-01re
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P11ARE

Quant Time: Jul 25 14:13:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	16596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	24746	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	21779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	8879	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	8607	45.511	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.020%
35) Dibromofluoromethane	7.715	113	6867	43.593	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.180%
50) Toluene-d8	10.178	98	28531	49.126	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.260%
62) 4-Bromofluorobenzene	12.477	95	8942	40.155	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
Data File : VY014792.D
Acq On : 25 Jul 2023 11:24
Operator : SY/MD
Sample : 03719-01re
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P11ARE

Quant Time: Jul 25 14:13:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

