

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
 Data File : VY014804.D
 Acq On : 25 Jul 2023 16:09
 Operator : SY/MD
 Sample : 03742-05
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P33B

Quant Time: Jul 26 01:19:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	22752	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	35127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	28472	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	11071	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	10260	39.573	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.140%
35) Dibromofluoromethane	7.716	113	9691	43.339	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.680%
50) Toluene-d8	10.179	98	40063	48.596	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.200%
62) 4-Bromofluorobenzene	12.483	95	11708	37.038	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	74.080%
Target Compounds						
16) Acetone	3.942	43	1743	21.215	ug/l	Qvalue # 61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
Data File : VY014804.D
Acq On : 25 Jul 2023 16:09
Operator : SY/MD
Sample : 03742-05
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P33B

Quant Time: Jul 26 01:19:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

