

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
 Data File : VY014810.D
 Acq On : 25 Jul 2023 18:31
 Operator : SY/MD
 Sample : 03733-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AU-6-072423

Quant Time: Jul 26 01:21:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	165009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	288740	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	277430	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	131941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	105088	55.888	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.780%
35) Dibromofluoromethane	7.716	113	91497	49.779	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.560%
50) Toluene-d8	10.179	98	345652	51.007	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.020%
62) 4-Bromofluorobenzene	12.477	95	128748	49.550	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.100%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.117	105	9162	1.082	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
Data File : VY014810.D
Acq On : 25 Jul 2023 18:31
Operator : SY/MD
Sample : 03733-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AU-6-072423

Quant Time: Jul 26 01:21:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

