

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
 Data File : VY014811.D
 Acq On : 25 Jul 2023 18:55
 Operator : SY/MD
 Sample : 03735-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 1029

Quant Time: Jul 26 01:22:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	165001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	289894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	277898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124055	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	102552	54.542	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.080%	
35) Dibromofluoromethane	7.716	113	92752	50.261	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.520%	
50) Toluene-d8	10.179	98	346884	50.986	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.980%	
62) 4-Bromofluorobenzene	12.483	95	126783	48.599	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 97.200%	
Target Compounds						Qvalue
11) Tert butyl alcohol	4.960	59	38946	88.410	ug/l #	82
52) Toluene	10.246	92	17888	3.562	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
Data File : VY014811.D
Acq On : 25 Jul 2023 18:55
Operator : SY/MD
Sample : 03735-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
1029

Quant Time: Jul 26 01:22:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

