

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
 Data File : VY014826.D
 Acq On : 26 Jul 2023 01:35
 Operator : SY/MD
 Sample : 03741-07
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P25B

Quant Time: Jul 26 03:35:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	165593	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	296716	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.495	117	282442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	131388	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	107901	57.181	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.360%
35) Dibromofluoromethane	7.728	113	94357	49.955	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.920%
50) Toluene-d8	10.184	98	353465	50.758	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.520%
62) 4-Bromofluorobenzene	12.483	95	129655	48.557	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
Data File : VY014826.D
Acq On : 26 Jul 2023 01:35
Operator : SY/MD
Sample : 03741-07
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P25B

Quant Time: Jul 26 03:35:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

