

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
 Data File : VY014829.D
 Acq On : 26 Jul 2023 02:45
 Operator : SY/MD
 Sample : 03741-10
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P27A

Quant Time: Jul 26 04:25:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	145790	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	248989	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.495	117	231127	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	109126	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	85455	51.438	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.880%
35) Dibromofluoromethane	7.728	113	77790	49.079	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.160%
50) Toluene-d8	10.185	98	297339	50.883	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.760%
62) 4-Bromofluorobenzene	12.483	95	107256	47.868	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
Data File : VY014829.D
Acq On : 26 Jul 2023 02:45
Operator : SY/MD
Sample : 03741-10
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P27A

Quant Time: Jul 26 04:25:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

