

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
 Data File : VY014832.D
 Acq On : 26 Jul 2023 03:55
 Operator : SY/MD
 Sample : 03741-13
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P6B

Quant Time: Jul 26 04:26:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	165768	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	295799	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.495	117	281708	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	129645	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	107510	56.914	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.820%
35) Dibromofluoromethane	7.728	113	92786	49.276	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.560%
50) Toluene-d8	10.185	98	352907	50.835	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.680%
62) 4-Bromofluorobenzene	12.483	95	128183	48.155	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.300%
Target Compounds						
16) Acetone	3.966	43	75425	126.003	ug/l	93
17) Carbon Disulfide	4.210	76	12423	2.658	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
Data File : VY014832.D
Acq On : 26 Jul 2023 03:55
Operator : SY/MD
Sample : 03741-13
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P6B

Quant Time: Jul 26 04:26:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

