



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072523\
Data File : VY014815.D
Acq On : 25 Jul 2023 20:50
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Title : SW846 8260
Last Update : Fri Jul 14 02:09:50 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	20498	PASS
75	95	30	60	52.3	55955	PASS
95	95	100	100	100.0	106976	PASS
96	95	5	9	6.6	7008	PASS
173	174	0.00	2	1.0	898	PASS
174	95	50	100	82.8	88592	PASS
175	174	5	9	7.4	6584	PASS
176	174	95	101	96.4	85405	PASS
177	176	5	9	6.5	5524	PASS