

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072524\
 Data File : VY018921.D
 Acq On : 25 Jul 2024 19:49
 Operator : SY/MD
 Sample : P3352-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL-COMP

Quant Time: Jul 26 03:02:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 09 14:28:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	33124	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.715	114	57808	50.000	ug/l	0.04
63) Chlorobenzene-d5	11.562	117	53011	50.000	ug/l	0.08
72) 1,4-Dichlorobenzene-d4	13.538	152	20519	50.000	ug/l	0.12
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	16133	49.496	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.000%
35) Dibromofluoromethane	7.722	113	18426	50.892	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.780%
50) Toluene-d8	10.233	98	64659	47.137	ug/l	0.06
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.280%
62) 4-Bromofluorobenzene	12.574	95	19602	41.845	ug/l	0.10
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.680%
Target Compounds						
20) Methylene Chloride	4.680	84	10466	28.632	ug/l	Qvalue # 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072524\
Data File : VY018921.D
Acq On : 25 Jul 2024 19:49
Operator : SY/MD
Sample : P3352-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL-COMP

Quant Time: Jul 26 03:02:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 09 14:28:51 2024
Response via : Initial Calibration

