

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072621\
 Data File : VY005455.D
 Acq On : 26 Jul 2021 19:17
 Operator : SY/MD
 Sample : M3159-09
 Misc : 2.78G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20210467-53

Quant Time: Jul 27 03:12:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	195811	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	328345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	324203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	162907	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	123661	59.114	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 118.220%	
35) Dibromofluoromethane	7.728	113	97012	51.170	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.340%	
50) Toluene-d8	10.185	98	416933	51.886	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.780%	
62) 4-Bromofluorobenzene	12.483	95	144979	54.666	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 109.340%	
Target Compounds						Qvalue
16) Acetone	3.960	43	29039	60.785	ug/l	96
17) Carbon Disulfide	4.204	76	55166	8.911	ug/l	99
25) 2-Butanone	7.002	43	8931	12.484	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072621\
Data File : VY005455.D
Acq On : 26 Jul 2021 19:17
Operator : SY/MD
Sample : M3159-09
Misc : 2.78G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20210467-53

Quant Time: Jul 27 03:12:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:20:14 2021
Response via : Initial Calibration

