

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072621\
 Data File : VY005457.D
 Acq On : 26 Jul 2021 20:04
 Operator : SY/MD
 Sample : M3152-07
 Misc : 7.07G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COLL-HOLE-A

Quant Time: Jul 27 03:13:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:20:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	227843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	345392	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	355990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	186014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	155687	63.960	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.920%	
35) Dibromofluoromethane	7.728	113	121670	61.008	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 122.020%	
50) Toluene-d8	10.185	98	440429	52.105	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.220%	
62) 4-Bromofluorobenzene	12.483	95	159328	57.112	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 114.220%	
Target Compounds						
16) Acetone	3.966	43	6063	10.907	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072621\
Data File : VY005457.D
Acq On : 26 Jul 2021 20:04
Operator : SY/MD
Sample : M3152-07
Misc : 7.07G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COLL-HOLE-A

Quant Time: Jul 27 03:13:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:20:14 2021
Response via : Initial Calibration

