



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072621\
Data File : VY005439.D
Acq On : 26 Jul 2021 09:31
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y072221S.M
Title : SW846 8260
Last Update : Fri Jul 23 06:20:14 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	19672	PASS
75	95	30	60	52.0	52251	PASS
95	95	100	100	100.0	100475	PASS
96	95	5	9	6.9	6924	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	89.0	89464	PASS
175	174	5	9	8.0	7149	PASS
176	174	95	101	96.2	86024	PASS
177	176	5	9	6.8	5888	PASS