

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072623\
 Data File : VY014847.D
 Acq On : 26 Jul 2023 14:19
 Operator : SY/MD
 Sample : 03717-01
 Misc : 6.20g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-1

Quant Time: Jul 26 14:37:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	162930	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	286310	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	275591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	133815	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	103128	55.545	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.100%
35) Dibromofluoromethane	7.716	113	91860	50.401	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.800%
50) Toluene-d8	10.173	98	343072	51.056	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.120%
62) 4-Bromofluorobenzene	12.477	95	129055	50.089	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	100.180%
Target Compounds						
						Qvalue
16) Acetone	3.954	43	4845	8.235	ug/l	94
25) 2-Butanone	6.984	43	7041	8.180	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072623\
Data File : VY014847.D
Acq On : 26 Jul 2023 14:19
Operator : SY/MD
Sample : 03717-01
Misc : 6.20g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-1

Quant Time: Jul 26 14:37:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

