

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072623\
 Data File : VY014850.D
 Acq On : 26 Jul 2023 15:29
 Operator : SY/MD
 Sample : 03741-03
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B-P12B

Quant Time: Jul 26 16:13:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 14 02:09:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	123801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	217170	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	207807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	98393	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	79586	56.414	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.820%
35) Dibromofluoromethane	7.716	113	69713	50.427	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.860%
50) Toluene-d8	10.179	98	257710	50.563	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.120%
62) 4-Bromofluorobenzene	12.477	95	95345	48.787	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.580%
Target Compounds						
16) Acetone	3.942	43	4438	9.927	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY072623\
Data File : VY014850.D
Acq On : 26 Jul 2023 15:29
Operator : SY/MD
Sample : 03741-03
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B-P12B

Quant Time: Jul 26 16:13:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y071323S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 14 02:09:50 2023
Response via : Initial Calibration

